

# EC Declaration of Conformity

In Accordance with EN ISO/IEC 17050-1:2010

Manufacturer: Nordson EFD LLC  
Address: 40 Catamore Boulevard  
East Providence, RI 02914 USA

Type of Equipment: Fluid Dispenser(s)  
Product Name: Performus Dispenser Models X15 / X100  
Model & Part Number:  Performus X15, 7363257  
 Performus X100, 7363256  
 Performus B10, 7364770

Manufacture Date: \_\_\_\_\_ Serial Number: \_\_\_\_\_

The above listed product(s) have been evaluated for conformity to the essential requirements of the below-listed directives:

WEEE Directive	2012/19/EU
RoHS Directive	2011/65/EU
EMC Directive	2014/30/EU
Low Voltage	2014/35/EU

With reference to the follow harmonized standards:

### Emissions Standards:

EN61000-6-3:2007+A1:2011	Conducted Emissions at AC Mains
EN61000-6-3:2007+A1:2011	Radiated Emissions
Subpart B-15.107	Conducted Emissions
Subpart B-15.109	Radiated Emissions
EN61000-3-2:2006 +A1:2009+	Harmonic Current Emissions
A2:2009	
EN61000-3-3:2008	Voltage Fluctuations Emissions

### Immunity Standards:

EN61000-4-2:1995+A1:1998+	Electrostatic Discharge (ESD) Immunity
A2:2010	
EN61000-4-3:2006+A1:2008+	RF Electromagnetic Field Immunity
A2:2010	
EN61000-4-4:2004+A1:2010	Fast Transient Immunity
EN61000-4-5:2006	Surges Immunity
EN61000-4-6:2013	Power-Frequency Magnetic Field Immunity
EN61000-4-11:2004	Voltage Dips/Voltage Interruptions Immunity

### Safety Standards:

IEC 61010-1:2010 Third Edition  
Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use,  
Part 1: General Requirements. This product conforms to UL STD 61010-1 and certified to CSA  
STD C22.2 # 61010-1.

Nordson EFD LLC



Sr. Quality Assurance Manager  
Place: East Providence, RI USA  
Date: April 6, 2020

Authorized Representative:

European Technical and Quality Manager  
Nordson Deutschland GmbH  
Nordson EFD Branch Office  
Raiffeisenalle 12b  
82041 Oberhaching Germany