



## Vector pull

The vector pull test was initially introduced to apply a programmable angled pull to leads on MBGA devices in order to remove the possibility of the pull hook sliding close to the neck of the bond.

The Nordson DAGE vector pull test negates the need for variable angle work holder with the use of software controlled programmable angle pull testing.